



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Standard Scale and Supply Company

25421 Glendale Avenue, Redford, MI 48239

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2005

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):

Calibration of Weigh Scales, Weight-Based Parts Counting Systems, Weight-Based Force Measurements Devices, Balances and Vehicle Weighing Systems
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

<i>Initial Accreditation Date:</i>	<i>Issue Date:</i>	<i>Accreditation No.:</i>	<i>Certificate No.:</i>
February 5, 2010	April 3, 2012	59180	L12-44

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjlab.com



Certificate of Accreditation: Supplement

Standard Scale and Supply Company

25421 Glendale Avenue, Redford, MI 48239

John Bowman Phone: 313-255-6700

Accreditation is granted to the facility to perform the following calibrations:

Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Scales	0.001 lb to 10 lb	$(1.0 \times 10^{-3} + 3.7 \times 10^{-5}Wt)$ lb	NIST Class F Weights
	0.01 lb to 100 lb	$(1.0 \times 10^{-2} + 1.14 \times 10^{-4}Wt)$ lb	
	0.1 lb to 1 000 lb	$(1.0 \times 10^{-1} + 1.15 \times 10^{-4}Wt)$ lb	
	1 lb to 10 000 lb	$(1.0 + 1.0 \times 10^{-4}Wt)$ lb	
	1 000 lb to 25 000 lb	23 lb	
	5 g to 50 kg	$(5.0 + 4.1 \times 10^{-5}Wt)$ g	
Industrial Balances	0.01 g to 10 000 g	$(1.0 \times 10^{-2} + 1.0 \times 10^{-4}Wt)$ g	ASTM Class 1 Weights
Analytical Balance	1 g to 400 g	$(1.41 \times 10^{-4} + 2.0 \times 10^{-6}Wt)$ g	
Force Testers	0.1 lb to 1 000 lb	$(1.0 \times 10^{-1} + 1.15 \times 10^{-4}Wt)$ lb	NIST Class F Weights

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represent the smallest measurement uncertainties attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The term Wt represents weight in pounds or grams (including SI multiple and submultiple units) appropriate to the uncertainty statement.



Perry Johnson Laboratory Accreditation, Inc.



May 9, 2012

Mr. John Bowman
Standard Scale and Supply Company
25421 Glendale Avenue
Redford, MI 48239

Dear Mr. Bowman:

This letter is to confirm that you have successfully completed your reaccreditation assessment. A certificate has now been granted and posted on our website. As you are aware, PJLA will no longer be issuing expiration dates on our certificates. Your certificate # **L12-44** will remain valid as long as you continue to maintain your annual assessments and reaccreditation assessments as stated in your customer agreement with PJLA. At this time, we have confirmed that your annual assessments will be conducted during the month of **November** each calendar year. This will include an interim surveillance assessment and a full system reassessment to be completed by **November 2014**. Once your reassessment is conducted and approved by our accreditation committee a revised status letter will be provided to you. Please allow PJLA at least 120 days from your assessment due date to issue this letter.

Please feel free to release this letter to any interested parties as confirmation of your certificate validity. Also, please remind them that your certificate is posted on our website at all times. Any changes in regards to your accreditation status will be reflected on our website.

We would like to thank you for your patronage over the past years and look forward to continuously serving your accreditation needs in the future. If we can assist you any further, please feel free to contact us at any time.

Sincerely,

Tracy Szerszen
President/Operations Manager